

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C | 30.363 °C | 30.0 °C | -0.363 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 %HR | 30.161 %HR | 32.9 %HR | 2.739 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.376 %HR | 49.6 %HR | -0.776 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 65 %HR | 65.335 %HR | 63.6 %HR | -1.735 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |