

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
15 °C	14.790 °C	15.0 °C	0.210 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.878 °C	25.1 °C	0.222 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.887 °C	32.1 °C	0.213 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	19.578 %HR	23.5 %HR	3.922 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.090 %HR	52.4 %HR	2.310 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.908 %HR	69.8 %HR	-1.108 %HR	± 5.0 %HR	± 1.9E+00 %HR