

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.105 °C	20.3 °C	0.195 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.751 °C	29.9 °C	0.149 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.977 %HR	52.1 %HR	2.123 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	71.685 %HR	69.3 %HR	-2.385 %HR	± 3.0 %HR	± 1.9E+00 %HR