

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
20 °C	20.035 °C	19.6 °C	-0.435 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	27.957 °C	27.5 °C	-0.457 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.491 °C	40.0 °C	-0.491 °C	± 0.5 °C	± 4.0E-01 °C