

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
-25 °C	-25.260 °C	-25.2 °C	0.060 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.271 °C	-0.2 °C	0.071 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.841 °C	8.0 °C	0.159 °C	± 0.5 °C	± 4.0E-01 °C