

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.150 °C	-25.30 °C	-0.150 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	0.170 °C	0.10 °C	-0.070 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.978 °C	25.30 °C	0.322 °C	± 0.5 °C	± 4.0E-01 °C
50 °C	49.706 °C	49.90 °C	0.194 °C	± 0.5 °C	± 4.0E-01 °C
80 °C	79.960 °C	80.20 °C	0.240 °C	± 0.5 °C	± 1.1E+00 °C
30 %HR	29.793 %HR	30.90 %HR	1.107 %HR	± 3.0 %HR	± 1.9E+00 %HR
40 %HR	41.419 %HR	43.40 %HR	1.981 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.468 %HR	51.70 %HR	1.232 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.161 %HR	68.90 %HR	-1.261 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	81.736 %HR	79.80 %HR	-1.936 %HR	± 3.0 %HR	± 1.9E+00 %HR