

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-25.5 °C	-0.411 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.310 °C	0.0 °C	-0.310 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.194 °C	4.9 °C	-0.294 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.107 °C	9.8 °C	-0.307 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.1 °C	-0.149 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.163 °C	29.8 °C	-0.363 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.641 °C	40.8 °C	0.159 °C	± 0.45 °C	± 4.0E-01 °C

