

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.062 °C	29.3 °C	-0.762 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	30.0 %HR	-0.161 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.376 %HR	46.0 %HR	-4.376 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	69.0 %HR	-1.652 %HR	± 5.0 %HR	± 1.9E+00 %HR