

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
3600.362 s	3601.00 s	0.638 s	0.018%	± 0.005 %	± 6.3E-02 s
3600.348 s	3601.00 s	0.652 s	0.018%	± 0.005 %	± 6.3E-02 s
3600.486 s	3601.00 s	0.514 s	0.014%	± 0.005 %	± 6.3E-02 s