

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.060 °C	-0.3 °C	-0.360 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.274 °C	5.1 °C	-0.174 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.238 °C	10.3 °C	0.062 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.220 %HR	32.2 %HR	2.980 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.118 %HR	52.0 %HR	2.882 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.897 %HR	71.2 %HR	0.303 %HR	± 3.0 %HR	± 1.9E+00 %HR