

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
49.98 mmHg	50 mmHg	0.02 mmHg	± 3 mmHg	±	6.6E-01 mmHg
99.98 mmHg	100 mmHg	0.02 mmHg	± 3 mmHg	±	6.5E-01 mmHg
149.98 mmHg	150 mmHg	0.02 mmHg	± 3 mmHg	±	6.6E-01 mmHg
199.99 mmHg	200 mmHg	0.01 mmHg	± 3 mmHg	±	6.8E-01 mmHg
249.99 mmHg	250 mmHg	0.01 mmHg	± 3 mmHg	±	7.5E-01 mmHg
299.99 mmHg	300 mmHg	0.01 mmHg	± 3 mmHg	±	7.5E-01 mmHg