

| Tipo de Medida (<i>Measurement Type</i>) | Patrón (<i>Standard</i>) | UBP (<i>UT</i>) | Error (<i>Error</i>) | E.M.P (<i>Tolerance</i>) | Incertidumbre (<i>Uncertainty</i>) |
|---|-----------------------------------|--------------------------|-------------------------------|-----------------------------------|---|
| Profundidad / (Depth) | 1 mm | 0.99 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 6.1\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 2 mm | 1.49 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 5.8\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 2 mm | 1.99 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 6.1\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 3 mm | 2.99 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 6.1\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 5 mm | 4.99 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 6.1\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 10 mm | 9.98 mm | -20 μm | $\pm 30 \mu\text{m}$ | $\pm 6.1\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 15 mm | 14.98 mm | -20 μm | $\pm 30 \mu\text{m}$ | $\pm 6.1\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 20 mm | 19.98 mm | -20 μm | $\pm 30 \mu\text{m}$ | $\pm 6.1\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 25 mm | 24.97 mm | -30 μm | $\pm 30 \mu\text{m}$ | $\pm 6.1\text{E}+00 \mu\text{m}$ |