

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.042 °C	29.9 °C	-0.142 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	35.0 %HR	4.634 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	54.0 %HR	3.645 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	75.0 %HR	4.348 %HR	± 5.0 %HR	± 1.9E+00 %HR