

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
8 °C	8.002 °C	8.0 °C	-0.002 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.092 °C	30.4 °C	0.308 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	29.8 %HR	-0.668 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	50.7 %HR	0.672 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	70.0 %HR	-0.243 %HR	± 3.0 %HR	± 1.9E+00 %HR