

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	15.1 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.119 °C	24.9 °C	-0.219 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.012 °C	29.6 °C	-0.412 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	57.9 %HR	-2.323 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.131 %HR	63.2 %HR	-1.931 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.2 %HR	-2.043 %HR	± 3.0 %HR	± 1.9E+00 %HR