

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
49.50 mmHg	50 mmHg	0.50 mmHg	± 5 mmHg	±	6.6E-01 mmHg
99.60 mmHg	100 mmHg	0.40 mmHg	± 5 mmHg	±	6.5E-01 mmHg
149.60 mmHg	150 mmHg	0.40 mmHg	± 5 mmHg	±	6.6E-01 mmHg
199.50 mmHg	200 mmHg	0.50 mmHg	± 5 mmHg	±	6.8E-01 mmHg
249.60 mmHg	250 mmHg	0.40 mmHg	± 5 mmHg	±	7.5E-01 mmHg
299.70 mmHg	300 mmHg	0.30 mmHg	± 5 mmHg	±	7.5E-01 mmHg