

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
3600.467 s	3601 s	0.123 s	0.0034%	± 0.01%	± 5.6E-02 s
3600.235 s	3600.360 s	0.125 s	0.0035%	± 0.01%	± 5.6E-02 s
3600.241 s	3600.37 s	0.129 s	0.0036%	± 0.01%	± 5.6E-02 s