

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.0625 in	0.063 in	0.0005 in	± 0.002 in	±	5.8E-04 in
0.1 in	0.101 in	0.001 in	± 0.002 in	±	5.8E-04 in
0.125 in	0.126 in	0.001 in	± 0.002 in	±	5.8E-04 in
0.1625 in	0.164 in	0.0015 in	± 0.002 in	±	5.9E-04 in
0.2 in	0.201 in	0.001 in	± 0.002 in	±	5.9E-04 in
0.25 in	0.251 in	0.001 in	± 0.002 in	±	6.0E-04 in
0.2625 in	0.263 in	0.0005 in	± 0.002 in	±	6.1E-04 in