

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.985 °C	2.1 °C	0.115 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.194 °C	5.4 °C	0.206 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.002 °C	8.2 °C	0.198 °C	± 0.5 °C	± 4.0E-01 °C