

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.130 °C	-25.2 °C	-0.070 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.090 °C	-0.2 °C	-0.290 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.224 °C	4.9 °C	-0.324 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.017 °C	9.6 °C	-0.417 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.299 °C	24.9 °C	-0.399 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.514 °C	30.4 °C	-0.114 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.501 °C	40.2 °C	-0.301 °C	± 0.45 °C	± 4.0E-01 °C

