

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.130 °C	-24.7 °C	0.430 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.090 °C	0.4 °C	0.310 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.224 °C	5.5 °C	0.276 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.017 °C	10.3 °C	0.283 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.299 °C	25.6 °C	0.301 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.514 °C	30.8 °C	0.286 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.501 °C	40.9 °C	0.399 °C	± 0.45 °C	± 4.0E-01 °C

