

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.130 °C	-24.6 °C	0.530 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.090 °C	0.5 °C	0.410 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.224 °C	5.6 °C	0.376 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.017 °C	10.4 °C	0.383 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.299 °C	25.6 °C	0.301 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.524 °C	31.0 °C	0.476 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.501 °C	40.8 °C	0.299 °C	± 0.5 °C	± 4.0E-01 °C

