

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.532 °C	15.60 °C	0.068 °C	± 0.1 °C	± 4.0E-01 °C
20 °C	20.045 °C	20.12 °C	0.075 °C	± 0.1 °C	± 4.0E-01 °C
30 °C	29.992 °C	29.95 °C	-0.042 °C	± 0.1 °C	± 4.0E-01 °C
30 %HR	30.907 %HR	32.2 %HR	1.293 %HR	± 1.5 %HR	± 1.9E+00 %HR
50 %HR	49.128 %HR	49.1 %HR	-0.028 %HR	± 1.5 %HR	± 1.9E+00 %HR
70 %HR	69.711 %HR	69.2 %HR	-0.511 %HR	± 1.5 %HR	± 1.9E+00 %HR