

Certificado No. (Certificate #):

65399

Equipo (Instrument): Medidor de Espesores

Fecha de Calibración (Calibration Date):

2026-01-29

Marca (Brand):

R&D Instruments

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
50 µm	50.2 µm	0.2 µm	±(3%+2µm)	± 3.3E+00 µm
100 µm	97.9 µm	-2.1 µm	±(3%+2µm)	± 2.1E+00 µm
150 µm	148.8 µm	-1.3 µm	±(3%+2µm)	± 2.0E+00 µm
248 µm	247.0 µm	-1.0 µm	±(3%+2µm)	± 2.3E+00 µm
348 µm	347.3 µm	-0.8 µm	±(3%+2µm)	± 2.6E+00 µm
495 µm	495.3 µm	0.3 µm	±(3%+2µm)	± 4.5E+00 µm
743 µm	752.0 µm	9.0 µm	±(3%+2µm)	± 2.3E+00 µm
1,001 µm	997.0 µm	-4.0 µm	±(3%+2µm)	± 5.0E+00 µm
1,249 µm	1,256.0 µm	7.0 µm	±(3%+2µm)	± 3.1E+00 µm