

**Clase de Exactitud (Accuracy Class):** Clase II (Class II)

---

<b>Valor UBP</b>	<b>Valor Patrón</b>	<b>Error</b>	<b>EMP</b>	<b>Incertidumbre</b>	<b>Valor UBP</b>	<b>Valor Patrón</b>	<b>Error</b>	<b>EMP</b>	<b>Incertidumbre</b>
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04942	0.580	± 0.6	± 4.3E-01					mm
0 a 0.5	0.49923	0.770	± 0.8	± 4.3E-01					mm
0 a 1	0.99922	0.780	± 1	± 4.3E-01					mm
0 a 2	1.99927	0.730	± 1.4	± 4.7E-01					mm
0 a 4	3.99937	0.630	± 2.2	± 4.7E-01					mm
0 a 5	4.99941	0.590	± 2.6	± 4.7E-01					mm
0 a 6	5.99945	0.550	± 3	± 3.4E-01					mm
0 a 7	6.99950	0.500	± 3.4	± 3.4E-01					mm
0 a 8	7.99955	0.450	± 3.8	± 3.4E-01					mm
0 a 10	9.99963	0.370	± 4.6	± 3.4E-01					mm
0 a 11	10.99967	0.330	± 5	± 3.4E-01					mm
0 a 12	11.99971	0.290	± 5.4	± 3.4E-01					mm
0 a 13	12.99975	0.250	± 5.8	± 3.4E-01					mm
0 a 14	13.99981	0.190	± 6.2	± 3.4E-01					mm
0 a 15	14.99985	0.150	± 6.6	± 3.4E-01					mm

---