

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
18 °C	18.099 °C	17.8 °C	-0.299 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.864 °C	19.6 °C	-0.264 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.758 °C	24.8 °C	0.042 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	27.666 °C	27.7 °C	0.034 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.312 %HR	40.9 %HR	1.588 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.905 %HR	52.9 %HR	2.995 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	59.660 %HR	63.3 %HR	3.640 %HR	± 3.0 %HR	± 1.9E+00 %HR