

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
18 °C	18.099 °C	17.9 °C	-0.199 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.864 °C	20.0 °C	0.136 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.758 °C	25.1 °C	0.342 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	27.666 °C	27.9 °C	0.234 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.312 %HR	43.0 %HR	3.688 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.905 %HR	56.4 %HR	6.495 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	59.660 %HR	68.1 %HR	8.440 %HR	± 3.0 %HR	± 1.9E+00 %HR