

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
21 °C	21.188 °C	21.4 °C	0.212 °C	± 0.5 °C	± 4.0E-01 °C
23 °C	23.454 °C	23.9 °C	0.446 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.758 °C	25.3 °C	0.542 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	27.666 °C	27.9 °C	0.234 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.312 %HR	37.7 %HR	-1.612 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.905 %HR	52.9 %HR	2.995 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	59.660 %HR	65.8 %HR	6.140 %HR	± 3.0 %HR	± 1.9E+00 %HR