

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
8 °C	7.661 °C	7 °C	-0.661 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.383 °C	19 °C	-0.383 °C	± 1.0 °C	± 4.0E-01 °C
36 °C	35.828 °C	35 °C	-0.828 °C	± 1.0 °C	± 4.0E-01 °C
45 %HR	44.957 %HR	43 %HR	-1.957 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.752 %HR	70 %HR	0.248 %HR	± 5.0 %HR	± 1.9E+00 %HR