

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.772 °C	29.8 °C	0.028 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.253 %HR	35 %HR	4.747 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.905 %HR	46 %HR	-3.905 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.325 %HR	65 %HR	-5.325 %HR	± 5.0 %HR	± 1.9E+00 %HR