

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
10 °C	10.007 °C	9.7 °C	-0.307 °C	± 0.1 °C	± 4.0E-01 °C
20 °C	20.065 °C	19.8 °C	-0.265 °C	± 0.1 °C	± 4.0E-01 °C
30 °C	30.263 °C	30.0 °C	-0.263 °C	± 0.1 °C	± 4.0E-01 °C
40 °C	40.190 °C	39.8 °C	-0.390 °C	± 0.1 °C	± 4.0E-01 °C