

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	9.8 °C	-0.207 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.065 °C	19.9 °C	-0.165 °C	± 2.0 °C	± 4.0E-01 °C
30 °C	30.263 °C	30.0 °C	-0.263 °C	± 2.0 °C	± 4.0E-01 °C
40 °C	40.190 °C	40.0 °C	-0.190 °C	± 2.0 °C	± 4.0E-01 °C