

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 10 °C | 10.007 °C | 9.8 °C | -0.207 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 20 °C | 20.065 °C | 19.9 °C | -0.165 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 °C | 30.263 °C | 30.0 °C | -0.263 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 40 °C | 40.190 °C | 40.0 °C | -0.190 °C | ± 0.5 °C | ± 4.0E-01 °C |