

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	14.9 °C	-0.322 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.4 °C	0.051 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.664 °C	30.5 °C	-0.164 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	61.3 %HR	1.077 %HR	± 3.5 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	68.5 %HR	3.165 %HR	± 3.5 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	73.2 %HR	2.957 %HR	± 3.5 %HR	± 1.9E+00 %HR