

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
15 °C	15.222 °C	15.1 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.3 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.664 °C	30.3 °C	-0.364 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	60.7 %HR	0.477 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	66.8 %HR	1.465 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72.7 %HR	2.457 %HR	± 3.0 %HR	± 1.9E+00 %HR