

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	15.5 °C	0.278 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.5 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.664 °C	30.0 °C	-0.664 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	62.5 %HR	2.277 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	66.5 %HR	1.165 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	70.5 %HR	0.257 %HR	± 3.0 %HR	± 1.9E+00 %HR