

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	15.5 °C	0.278 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.6 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.664 °C	30.6 °C	-0.064 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	59.6 %HR	-0.623 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	66.2 %HR	0.865 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72.6 %HR	2.357 %HR	± 3.0 %HR	± 1.9E+00 %HR