

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
15 °C	15.222 °C	15.3 °C	0.078 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.5 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.664 °C	30.4 °C	-0.264 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	60.5 %HR	0.277 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	67.2 %HR	1.865 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72.5 %HR	2.257 %HR	± 3.0 %HR	± 1.9E+00 %HR