

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.020 °C	0.09 °C	0.112 °C	± 0.21 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.91 °C	-0.097 °C	± 0.21 °C	± 4.0E-01 °C
20 °C	20.035 °C	19.96 °C	-0.075 °C	± 0.21 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.96 °C	-0.102 °C	± 0.21 °C	± 4.0E-01 °C
40 °C	40.090 °C	40.23 °C	0.140 °C	± 0.21 °C	± 4.0E-01 °C
50 °C	50.118 °C	50.19 °C	0.072 °C	± 0.21 °C	± 1.1E+00 °C
20 %HR	20.345 %HR	22.30 %HR	1.955 %HR	± 3.50 %HR	± 1.9E+00 %HR
35 %HR	35.171 %HR	36.90 %HR	1.729 %HR	± 3.50 %HR	± 1.9E+00 %HR
50 %HR	50.130 %HR	52.30 %HR	2.170 %HR	± 3.50 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	78.30 %HR	2.740 %HR	± 3.50 %HR	± 1.9E+00 %HR
80 %HR	80.264 %HR	82.10 %HR	1.836 %HR	± 3.50 %HR	± 1.9E+00 %HR