

Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
10 Hz	0.800 g-pk	0.83 g-pk	0.030 g-pk	± 0.008 g-pk	±	1.7E-02 g-pk
30 Hz	1.000 g-pk	1.03 g-pk	0.030 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
50 Hz	1.000 g-pk	1.02 g-pk	0.020 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
100 Hz	1.000 g-pk	1.01 g-pk	0.010 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
500 Hz	1.000 g-pk	1.02 g-pk	0.020 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
1000 Hz	1.000 g-pk	0.96 g-pk	-0.040 g-pk	± 0.01 g-pk	±	2.2E-02 g-pk