

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.10000	0.26	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.14999	0.27	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.19998	0.27	± 0.6	± 4.3E-01 mm					
0 a 0.3	0.29998	0.27	± 0.7	± 4.3E-01 mm					
0 a 0.4	0.39998	0.28	± 0.7	± 4.3E-01 mm					
0 a 0.5	0.49999	0.27	± 0.8	± 4.3E-01 mm					
0 a 0.6	0.59999	0.28	± 0.8	± 4.3E-01 mm					
0 a 0.7	0.69998	0.29	± 0.8	± 4.3E-01 mm					
0 a 0.8	0.79998	0.30	± 0.9	± 4.3E-01 mm					
0 a 0.9	0.89997	0.30	± 0.9	± 4.3E-01 mm					
0 a 1	0.99997	0.31	± 1	± 4.3E-01 mm					
