

| <b>Patrón</b> ( <i>Standard</i> ) | <b>UBP</b> ( <i>UT</i> ) | <b>Error</b> ( <i>Error</i> ) | <b>Exactitud</b> ( <i>Accuracy</i> ) | <b>Incertidumbre</b> ( <i>Uncertainty</i> ) |
|-----------------------------------|--------------------------|-------------------------------|--------------------------------------|---|
| 0.4986 ml                         | 0.5000 ml                | 0.0014 ml                     | ± 0.01 ml                            | ± 5.3E-02 µl                                |
| 2.4933 ml                         | 2.5000 ml                | 0.0067 ml                     | ± 0.01 ml                            | ± 1.1E-01 µl                                |
| 5.0085 ml                         | 5.0000 ml                | -0.0085 ml                    | ± 0.01 ml                            | ± 1.7E-01 µl                                |