

Patrón (Standard)	UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2.4916 ml	2.5000 ml	0.0084 ml	± 0.03 ml	± 1.1E-01 µl
12.5080 ml	12.5000 ml	-0.0080 ml	± 0.03 ml	± 3.9E-01 µl
24.9606 ml	25.0000 ml	0.0394 ml	± 0.03 ml	± 7.8E-01 µl