

Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10.0114 ml	10.0000 ml	-0.0114 ml	± 0.05 ml	± 3.1E-01 µl
24.9597 ml	25.0000 ml	0.0403 ml	± 0.05 ml	± 7.8E-01 µl
49.9852 ml	50.0000 ml	0.0148 ml	± 0.05 ml	± 1.5E+00 µl