

<b>Patrón (Standard)</b>	<b>UBP (UT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
10.0002 ml	10.0000 ml	-0.0002 ml	± 0.05 ml	± 3.1E-01 µl
24.9236 ml	25.0000 ml	0.0764 ml	± 0.05 ml	± 7.8E-01 µl
49.9135 ml	50.0000 ml	0.0865 ml	± 0.05 ml	± 1.5E+00 µl