

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10.0002 ml	10.0000 ml	-0.0002 ml	± 0.05 ml	± 3.1E-01 µl
24.9795 ml	25.0000 ml	0.0205 ml	± 0.05 ml	± 7.8E-01 µl
49.9857 ml	50.0000 ml	0.0143 ml	± 0.05 ml	± 1.5E+00 µl