

Certificado No. (Certificate #):

65634

Equipo (Instrument): Calibrador Pie de Rey

Fecha de Calibración (Calibration Date):

2026-01-29

Marca (Brand):

-

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	1.000 in	1.000 in	0 in	± 0.0019685 in	± 4.7E-03 in
Exterior / (Exterior)	3.000 in	3.000 in	0 in	± 0.0019685 in	± 4.7E-03 in
Exterior / (Exterior)	5.000 in	5.000 in	0 in	± 0.0027559 in	± 4.7E-03 in
Exterior / (Exterior)	6.000 in	6.000 in	0 in	± 0.0027559 in	± 4.7E-03 in
Exterior / (Exterior)	7.000 in	7.000 in	0 in	± 0.0027559 in	± 4.7E-03 in
Interior / (Interior)	1.000 in	1.000 in	0 in	± 0.0019685 in	± 4.5E-03 in
Interior / (Interior)	6.000 in	6.000 in	0 in	± 0.0027559 in	± 4.5E-03 in
Profundidad / (Depth)	2.000 in	2.000 in	0 in	± 0.0019685 in	± 4.6E-03 in
Profundidad / (Depth)	4.000 in	4.000 in	0 in	± 0.0027559 in	± 4.6E-03 in