

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
49.90 mmHg	50 mmHg	0.10 mmHg	± 3 mmHg	±	6.6E-01 mmHg
100.20 mmHg	100 mmHg	-0.20 mmHg	± 3 mmHg	±	6.5E-01 mmHg
150.10 mmHg	150 mmHg	-0.10 mmHg	± 3 mmHg	±	6.6E-01 mmHg
200.20 mmHg	200 mmHg	-0.20 mmHg	± 3 mmHg	±	6.8E-01 mmHg
250.10 mmHg	250 mmHg	-0.10 mmHg	± 3 mmHg	±	7.5E-01 mmHg
299.90 mmHg	300 mmHg	0.10 mmHg	± 3 mmHg	±	7.5E-01 mmHg