

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
-10 °C	-10.249 °C	-10.3 °C	-0.051 °C	± 2.0 °C	± 4.0E-01 °C
2 °C	2.266 °C	2.3 °C	0.034 °C	± 2.0 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.5 °C	0.493 °C	± 2.0 °C	± 4.0E-01 °C