

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
-10 °C	-10.249 °C	-10.3 °C	-0.051 °C	± 2.0 °C	± 4.0E-01 °C
2 °C	2.266 °C	2.3 °C	0.034 °C	± 2.0 °C	± 4.0E-01 °C
10 °C	10.007 °C	11.2 °C	1.193 °C	± 2.0 °C	± 4.0E-01 °C